

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.809 °C	10.6 °C	-0.209 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.870 °C	32.5 °C	-0.370 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.683 %HR	35.0 %HR	-0.683 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.560 %HR	77.7 %HR	2.140 %HR	± 3.0 %HR	± 1.9E+00 %HR