

Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	E.M.P (<i>Tolerance</i>)	Incertidumbre (<i>Uncertainty</i>)	
23.3 μm	23.2 μm	-0.1 μm	± 2.5 μm	±	5.8E-02 μm
51.6 μm	51.5 μm	-0.1 μm	± 2.5 μm	±	5.8E-02 μm
124.6 μm	124 μm	-0.6 μm	± 3.7 μm	±	6.1E-02 μm
250.3 μm	247 μm	-3.3 μm	± 7.5 μm	±	6.9E-02 μm
486.0 μm	483 μm	-3.0 μm	± 15 μm	±	9.4E-02 μm
963.0 μm	961 μm	-2.0 μm	± 29 μm	±	1.6E-01 μm