

| Nominal <i>(Nominal)</i> | Patrón <i>(Standard)</i> | UBP <i>(UUT)</i> | Error <i>(Error)</i> | Exactitud <i>(Accuracy)</i> | Incertidumbre <i>(Uncertainty)</i> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 0 °C                     | -0.201 °C                | -0.5 °C          | -0.299 °C            | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 10 °C                    | 9.997 °C                 | 9.5 °C           | -0.497 °C            | ± 0.5 °C                    | ± 4.0E-01 °C                       |