

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.201 °C	-0.5 °C	-0.299 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	9.997 °C	9.5 °C	-0.497 °C	± 0.5 °C	± 4.0E-01 °C