

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.062 °C	30.1 °C	0.038 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	33.5 %HR	2.930 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.141 %HR	51.3 %HR	1.159 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.448 %HR	69.4 %HR	-1.048 %HR	± 3.0 %HR	± 1.9E+00 %HR