

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.05	0.04904	0.960	± 1.2	± 4.3E-01 mm					
0 a 0.1	0.09906	0.940	± 1.2	± 4.3E-01 mm					
0 a 0.15	0.14902	0.980	± 1.3	± 4.3E-01 mm					
0 a 1	0.99908	0.920	± 2	± 4.3E-01 mm					
0 a 2	1.99912	0.880	± 2.8	± 4.8E-01 mm					
0 a 3	2.99912	0.880	± 3.6	± 4.8E-01 mm					
0 a 4	3.99920	0.800	± 4.4	± 4.8E-01 mm					
0 a 5	4.99931	0.690	± 5.2	± 4.8E-01 mm					
0 a 8	7.99944	0.560	± 7.6	± 3.6E-01 mm					
0 a 10	9.99951	0.490	± 9.2	± 3.6E-01 mm					
0 a 12	11.99984	0.160	± 10.8	± 3.6E-01 mm					
0 a 15	14.99975	0.250	± 13.2	± 3.6E-01 mm					
0 a 20	19.99993	0.070	± 17.2	± 3.6E-01 mm					
0 a 25	25.00014	-0.140	± 21.2	± 5.5E-01 mm					
0 a 30	30.00032	-0.320	± 25.2	± 5.5E-01 mm					