

Clase de Exactitud (Accuracy Class): Clase III (Class III)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.5	0.49906	0.940	± 1.6	± 4.3E-01 mm					
0 a 1	0.99910	0.900	± 2	± 4.3E-01 mm					
0 a 2	1.99917	0.830	± 2.8	± 4.8E-01 mm					
0 a 4	3.99924	0.760	± 4.4	± 4.8E-01 mm					
0 a 6	5.99933	0.670	± 6	± 3.6E-01 mm					
0 a 10	9.99953	0.470	± 9.2	± 3.6E-01 mm					
0 a 15	14.99973	0.270	± 13.2	± 3.6E-01 mm					
0 a 18	17.99986	0.140	± 15.6	± 3.6E-01 mm					
0 a 21	20.99998	0.020	± 18	± 5.5E-01 mm					
0 a 24	24.00008	-0.080	± 20.4	± 5.5E-01 mm					
0 a 30	30.00054	-0.540	± 25.2	± 5.5E-01 mm					
0 a 35	35.00050	-0.500	± 29.2	± 5.5E-01 mm					
0 a 40	40.00065	-0.650	± 33.2	± 5.5E-01 mm					
0 a 45	45.00082	-0.820	± 37.2	± 6.3E-01 mm					
0 a 50	50.00097	-0.970	± 41.2	± 6.3E-01 mm					