

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.343 °C	29.1 °C	-1.243 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.325 %HR	38 %HR	7.675 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.783 %HR	55 %HR	5.217 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.875 %HR	83 %HR	8.125 %HR	± 5.0 %HR	± 1.9E+00 %HR