

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1 mm	1.00 mm	0.0 µm	± 20 µm	±	5.8E+00 µm
2 mm	2.01 mm	10.0 µm	± 20 µm	±	5.8E+00 µm
3 mm	3.01 mm	10.0 µm	± 20 µm	±	5.8E+00 µm
4 mm	4.02 mm	20.0 µm	± 20 µm	±	5.8E+00 µm
5 mm	5.01 mm	10.0 µm	± 20 µm	±	5.8E+00 µm
6 mm	6.01 mm	10.0 µm	± 20 µm	±	5.8E+00 µm
7 mm	7.01 mm	10.0 µm	± 20 µm	±	5.8E+00 µm
8 mm	8.01 mm	10.0 µm	± 20 µm	±	5.8E+00 µm
10 mm	10.00 mm	0.0 µm	± 20 µm	±	5.8E+00 µm