

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.0946 µl	100 µl	-0.095 µl	± 0.1 µl	±	5.3E-02 µl
500.0220 µl	500 µl	-0.022 µl	± 0.1 µl	±	5.7E-02 µl
1000.044 µl	1000 µl	-0.044 µl	± 0.1 µl	±	7.0E-02 µl