

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
99.9702 µl	100 µl	0.030 µl	± 0.1 µl	±	5.3E-02 µl
500.0516 µl	500 µl	-0.052 µl	± 0.1 µl	±	6.0E-02 µl
999.943 µl	1000 µl	0.057 µl	± 0.1 µl	±	8.2E-02 µl