

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
9.9041 µl	10 µl	0.096 µl	± 0.1 µl	±	7.7E-02 µl
50.0720 µl	50 µl	-0.072 µl	± 0.1 µl	±	6.1E-02 µl
99.944 µl	100 µl	0.056 µl	± 0.1 µl	±	5.3E-02 µl