

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.0438 µl	100 µl	-0.044 µl	± 0.1 µl	±	5.3E-02 µl
499.9182 µl	500 µl	0.082 µl	± 0.1 µl	±	5.9E-02 µl
999.937 µl	1000 µl	0.063 µl	± 0.1 µl	±	7.8E-02 µl