

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	
m	m	mm	mm			m	m	mm	mm		
0 a 0.3	0.29997	0.030	± 0.7	±	4.3E-01 mm						
0 a 0.6	0.60014	-0.140	± 0.8	±	4.3E-01 mm						
0 a 0.9	0.90032	-0.320	± 0.9	±	4.3E-01 mm						
0 a 1.2	1.20033	-0.330	± 1	±	4.8E-01 mm						
0 a 1.5	1.50030	-0.300	± 1.2	±	4.8E-01 mm						
0 a 1.8	1.80033	-0.330	± 1.3	±	4.8E-01 mm						
0 a 2.1	2.10021	-0.210	± 1.4	±	4.8E-01 mm						
0 a 2.4	2.39970	0.300	± 1.5	±	4.8E-01 mm						
0 a 2.7	2.69945	0.550	± 1.6	±	4.8E-01 mm						
0 a 3	2.99951	0.490	± 1.8	±	4.8E-01 mm						