

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	29.992 °C	29.3 °C	-0.692 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.118 %HR	33 %HR	3.882 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.220 %HR	44 %HR	-5.220 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.251 %HR	64 %HR	-10.251 %HR	± 5.0 %HR	± 1.9E+00 %HR