

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.992 °C	29.3 °C	-0.692 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	29.118 %HR	32 %HR	2.882 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.220 %HR	45 %HR	-4.220 %HR	± 5.0 %HR	± 1.9E+00 %HR
75 %HR	74.251 %HR	69 %HR	-5.251 %HR	± 5.0 %HR	± 1.9E+00 %HR