

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.0157 ml	1.0 ml	-0.016 ml	± 0.6 ml	±	9.5E-02 µl
4.9799 ml	5.0 ml	0.020 ml	± 0.6 ml	±	3.0E-01 µl
9.9828 ml	10.0 ml	0.017 ml	± 0.6 ml	±	5.9E-01 µl