

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0.5042 ml	0.5 ml	-0.004 ml	± 0.6 ml	±	9.9E-02 µl
3.0096 ml	3.0 ml	-0.010 ml	± 0.6 ml	±	2.0E-01 µl
5.0090 ml	5.0 ml	-0.009 ml	± 0.6 ml	±	3.0E-01 µl