

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.0939 µl	100 µl	-0.094 µl	± 0.6 µl	±	5.3E-02 µl
499.9182 µl	500.0 µl	0.082 µl	± 0.6 µl	±	5.9E-02 µl
999.9367 µl	1000.0 µl	0.063 µl	± 0.6 µl	±	7.8E-02 µl