

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
100.059 µl	100 µl	-0.059 µl	± 0.6 µl	±	5.2E-02 µl
500.042 µl	500 µl	-0.042 µl	± 0.6 µl	±	5.3E-02 µl
1000.285 µl	1000 µl	-0.285 µl	± 0.6 µl	±	5.9E-02 µl