

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
25.02417 ml	25 ml	-0.02417 ml	± 0.03 ml	±	7.7E-01 µl