

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
2.0051 ml	2.0 ml	-0.005 ml	± 0.01 ml	±	9.8E-02 µl
5.0080 ml	5.0 ml	-0.008 ml	± 0.01 ml	±	1.7E-01 µl
10.0028 ml	10.0 ml	-0.003 ml	± 0.01 ml	±	3.2E-01 µl