

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.121 °C	15.0 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.2 °C	0.151 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.268 °C	32.2 °C	-0.068 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	19.4 %HR	-0.945 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.079 %HR	50.4 %HR	0.321 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	67.2 %HR	-2.941 %HR	± 3.0 %HR	± 1.9E+00 %HR