

| Nominal ( <i>Nominal</i> ) | Patrón ( <i>Standard</i> ) | UBP ( <i>UUT</i> ) | Error ( <i>Error</i> ) | Exactitud ( <i>Accuracy</i> ) | Incertidumbre ( <i>Uncertainty</i> ) |
|----------------------------|----------------------------|--------------------|------------------------|-------------------------------|--------------------------------------|
| 15 °C                      | 15.121 °C                  | 15.0 °C            | -0.121 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 25 °C                      | 25.049 °C                  | 24.8 °C            | -0.249 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 32 °C                      | 32.268 °C                  | 32.2 °C            | -0.068 °C              | ± 0.5 °C                      | ± 4.0E-01 °C                         |
| 20 %HR                     | 20.345 %HR                 | 19.4 %HR           | -0.945 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 50 %HR                     | 50.079 %HR                 | 50.6 %HR           | 0.521 %HR              | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |
| 70 %HR                     | 70.141 %HR                 | 69.4 %HR           | -0.741 %HR             | ± 3.0 %HR                     | ± 1.9E+00 %HR                        |