

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.121 °C	15.0 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	24.8 °C	-0.249 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.268 °C	32.2 °C	-0.068 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	19.4 %HR	-0.945 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.079 %HR	50.6 %HR	0.521 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	69.4 %HR	-0.741 %HR	± 5.0 %HR	± 1.9E+00 %HR