

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.121 °C	15.5 °C	0.379 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.049 °C	25.0 °C	-0.049 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.268 °C	32.7 °C	0.432 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.345 %HR	21.4 %HR	1.055 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.079 %HR	51.2 %HR	1.121 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.141 %HR	72.2 %HR	2.059 %HR	± 5.0 %HR	± 1.9E+00 %HR