

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
-25 °C	-25.260 °C	-24.6 °C	0.660 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.271 °C	0.4 °C	0.671 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.841 °C	8.6 °C	0.759 °C	± 0.5 °C	± 4.0E-01 °C