

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| -25 °C            | -25.260 °C        | -24.7 °C  | 0.560 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 0 °C              | -0.271 °C         | 0.6 °C    | 0.871 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |
| 8 °C              | 7.841 °C          | 8.9 °C    | 1.059 °C      | ± 0.5 °C             | ± 4.0E-01 °C                |