

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.260 °C	-24.5 °C	0.760 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.271 °C	0.4 °C	0.671 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.841 °C	8.5 °C	0.659 °C	± 0.5 °C	± 4.0E-01 °C