

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.260 °C	-24.6 °C	0.660 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.271 °C	0.7 °C	0.971 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.841 °C	8.9 °C	1.059 °C	± 0.5 °C	± 4.0E-01 °C