

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
0 mm	0.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
1.00 mm	1.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
2.0 mm	2.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
3 mm	3.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
5 mm	5.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
7 mm	7.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
10 mm	10.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
15 mm	15.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
20 mm	20.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
25 mm	25.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
30 mm	30.00 mm	0 µm	± 10 µm	±	5.8E+00 µm