

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
# 0 mm	0.00 mm	0 μm	± 10 μm	±	5.8E+00 μm
1.00 mm	1.00 mm	0 μm	± 10 μm	±	5.8E+00 μm
2.0 mm	2.00 mm	0 μm	± 10 μm	±	5.8E+00 μm
3 mm	3.00 mm	0 μm	± 10 μm	±	5.8E+00 μm
5 mm	5.00 mm	0 μm	± 10 μm	±	5.8E+00 μm
7 mm	7.00 mm	0 μm	± 10 μm	±	5.8E+00 μm
10 mm	10.00 mm	0 μm	± 10 μm	±	5.8E+00 μm
15 mm	15.00 mm	0 μm	± 10 μm	±	5.8E+00 μm
20 mm	20.00 mm	0 μm	± 10 μm	±	5.8E+00 μm
25 mm	25.00 mm	0 μm	± 10 μm	±	5.8E+00 μm
30 mm	30.00 mm	0 μm	± 10 μm	±	5.8E+00 μm