

Patrón <i>(Standard)</i>	UBP <i>(UUT)</i>	Error <i>(Error)</i>	E.M.P <i>(Tolerance)</i>	Incertidumbre <i>(Uncertainty)</i>	
1.005 in	1.005 in	0.000 in	± 50.25 in	±	5.8E-05 in
1.01 in	1.013 in	0.003 in	± 50.5 in	±	5.8E-05 in
1.05 in	1.051 in	0.001 in	± 52.5 in	±	5.8E-05 in
1.10 in	1.102 in	0.002 in	± 55. in	±	5.8E-05 in
1.30 in	1.300 in	0.000 in	± 65 in	±	5.8E-05 in