

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) |            |
|-------------------|-----------|---------------|-------------------|-----------------------------|------------|
| 52.65 μm          | 52.92 μm  | 0.27 μm       | ± 2.5 μm          | ±                           | 5.8E-01 μm |
| 155.23 μm         | 154.6 μm  | -0.63 μm      | ± 2.5 μm          | ±                           | 5.8E-01 μm |
| 257.12 μm         | 257.2 μm  | 0.08 μm       | ± 2.5 μm          | ±                           | 5.8E-01 μm |
| 511.72 μm         | 512.0 μm  | 0.28 μm       | ± 2.5 μm          | ±                           | 5.8E-01 μm |