

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
52.65 μm	52.58 μm	-0.07 μm	± 2.5 μm	±	5.8E-01 μm
155.23 μm	154.8 μm	-0.43 μm	± 2.5 μm	±	5.8E-01 μm
257.12 μm	256.6 μm	-0.52 μm	± 2.5 μm	±	5.8E-01 μm
511.72 μm	511.6 μm	-0.12 μm	± 2.5 μm	±	5.8E-01 μm