

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.082 °C	29.7 °C	-0.382 °C	± 1.5 °C	± 4.0E-01 °C
30 %HR	29.118 %HR	30.8 %HR	1.662 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	74.916 %HR	75.8 %HR	0.884 %HR	± 3.0 %HR	± 1.9E+00 %HR
Medición Punto de Rocío					
Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
5 °C	5.013 °C	3.5 °C	-1.513 °C	± 1.5 °C	± 4.0E-01 °C
25 °C	25.008 °C	24.4 °C	-0.608 °C	± 1.5 °C	± 4.0E-01 °C
Medición Temperatura Bulbo Humedo					
Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	20.426 °C	7.8 °C	-12.6 °C	± 1.5 °C	± 4.0E-01 °C
30 °C	30.022 °C	19.7 °C	-10.3 °C	± 1.5 °C	± 4.0E-01 °C