

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.822 °C	30.1 °C	0.278 °C	± 1.5 °C	± 4.0E-01 °C
30 %HR	29.118 %HR	33.4 %HR	4.282 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.110 %HR	70.6 %HR	-4.510 %HR	± 3.0 %HR	± 1.9E+00 %HR
Medición Punto de Rocío					
Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
5 °C	5.214 °C	4.9 °C	-0.334 °C	± 1.5 °C	± 4.0E-01 °C
25 °C	25.159 °C	23.6 °C	-1.559 °C	± 1.5 °C	± 4.0E-01 °C
Medición Temperatura Bulbo Humedo					
Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
20 °C	19.614 °C	15.3 °C	-4.3 °C	± 1.5 °C	± 4.0E-01 °C
30 °C	29.982 °C	24.6 °C	-5.4 °C	± 1.5 °C	± 4.0E-01 °C