

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1 mm	1.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
1.25 mm	1.25 mm	0 µm	± 10 µm	±	5.8E+00 µm
1.5 mm	1.49 mm	-10 µm	± 10 µm	±	5.8E+00 µm
2 mm	2.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
3 mm	2.99 mm	-10 µm	± 10 µm	±	5.8E+00 µm
5 mm	4.99 mm	-10 µm	± 10 µm	±	5.8E+00 µm
6 mm	6.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
8 mm	8.00 mm	0 µm	± 10 µm	±	5.8E+00 µm
10 mm	10.01 mm	10 µm	± 10 µm	±	5.8E+00 µm