

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre		Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	
m	m	mm	mm	(Uncertainty)		m	m	mm	mm	(Uncertainty)	
0 a 0.1	0.09997	0.030	± 0.2	±	4.3E-01 mm						
0 a 0.2	0.19999	0.010	± 0.2	±	4.3E-01 mm						
0 a 0.3	0.30001	-0.010	± 0.2	±	4.3E-01 mm						
0 a 0.4	0.39996	0.040	± 0.2	±	4.3E-01 mm						
0 a 0.5	0.49997	0.030	± 0.3	±	4.3E-01 mm						
0 a 0.6	0.59997	0.030	± 0.3	±	4.3E-01 mm						
0 a 0.7	0.69998	0.020	± 0.3	±	4.3E-01 mm						
0 a 0.8	0.79999	0.010	± 0.3	±	4.3E-01 mm						
0 a 0.9	0.90002	-0.020	± 0.3	±	4.3E-01 mm						
0 a 1	1.00005	-0.050	± 0.4	±	4.3E-01 mm						