

Patrón <i>(Standard)</i>	UBP <i>(UT)</i>	Error <i>(Error)</i>		Exactitud <i>(Accuracy)</i>	Incertidumbre <i>(Uncertainty)</i>	
3600.224 s	3600.21 s	-0.014 s	-0.0004%	± 0.01 %	±	5.6E-02 s
3600.068 s	3600.07 s	0.002 s	0.0001%	± 0.01 %	±	5.6E-02 s
3600.111 s	3600.1 s	-0.011 s	-0.0003%	± 0.01 %	±	5.6E-02 s