

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
in	in	in	in	(Uncertainty)	in	in	in	in	(Uncertainty)
0 a 0.25	0.250	0.00	± 0.01	± 3.2E-04 in					
0 a 0.5	0.500	0.00	± 0.3	± 3.2E-04 in					
0 a 0.75	0.750	0.00	± 0.3	± 3.2E-04 in					
0 a 1	1.000	0.00	± 0.4	± 3.2E-04 in					
0 a 1.25	1.250	0.00	± 0.4	± 3.9E-04 in					
0 a 1.5	1.500	0.00	± 0.5	± 3.9E-04 in					
0 a 1.75	1.750	0.00	± 0.5	± 3.9E-04 in					
0 a 2	2.000	0.00	± 0.6	± 3.9E-04 in					