

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	29.982 °C	30.1 °C	0.118 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.732 %HR	32.7 %HR	2.968 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.026 %HR	49.2 %HR	0.174 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	75.274 %HR	72.5 %HR	-2.774 %HR	± 3.0 %HR	± 1.9E+00 %HR