

Certificado No. (Certificate #):57691Equipo (Instrument):Reloj Indicador

Fecha de Calibración (Calibration Date):2025-06-04Marca (Brand):Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.0000 mm	1.00 mm	0 µm		±	5.8E+00 µm
1.0500 mm	1.05 mm	0 µm		±	5.8E+00 µm
3.0000 mm	3.00 mm	0 µm		±	5.8E+00 µm
5.0000 mm	5.00 mm	0 µm		±	5.8E+00 µm
7.0000 mm	7.00 mm	0 µm		±	5.8E+00 µm
8.0000 mm	8.00 mm	0 µm		±	5.8E+00 µm
10.0000 mm	10.00 mm	0 µm		±	5.8E+00 µm
12.0000 mm	12.00 mm	0 µm		±	5.8E+00 µm
13.0000 mm	13.00 mm	0 µm		±	5.8E+00 µm