

Certificado No. (Certificate #):57692Equipo (Instrument):Reloj Indicador

Fecha de Calibración (Calibration Date):2025-06-04Marca (Brand):Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
1.0000 mm	0.993 mm	-7 µm		±	5.8E-01 µm
1.0500 mm	1.044 mm	-6 µm		±	5.8E-01 µm
3.0000 mm	2.992 mm	-8 µm		±	5.8E-01 µm
5.0000 mm	4.991 mm	-9 µm		±	5.8E-01 µm
7.0000 mm	6.991 mm	-9 µm		±	5.8E-01 µm
8.0000 mm	7.991 mm	-9 µm		±	5.8E-01 µm
10.0000 mm	9.990 mm	-10 µm		±	5.9E-01 µm
12.0000 mm	11.994 mm	-6 µm		±	5.9E-01 µm
14.0000 mm	13.995 mm	-5 µm		±	6.0E-01 µm