

Certificado No. (Certificate #):		57693		Equipo (Instrument):		Micrómetro de Exterior			
Fecha de Calibración (Calibration Date):		2025-06-04		Marca (Brand):		Mitutoyo			
Patrón (Standard)		UBP (UUT)		Error (Error)		E.M.P (Tolerance)		Incertidumbre (Uncertainty)	
1.0000 mm		1.0010 mm		1.0000 µm		± 4 µm		± 3.9E+00 µm	
1.0500 mm		1.0502 mm		0.2000 µm		± 4 µm		± 4.1E+00 µm	
3.0000 mm		2.9999 mm		-0.1000 µm		± 4 µm		± 1.2E+01 µm	
5.0000 mm		5.0013 mm		1.3000 µm		± 4 µm		± 1.9E+01 µm	
7.0000 mm		7.0003 mm		0.3000 µm		± 4 µm		± 2.7E+01 µm	
10.0000 mm		10.0008 mm		0.8000 µm		± 4 µm		± 3.9E+01 µm	
13.0000 mm		13.0025 mm		2.5000 µm		± 4 µm		± 5.0E+01 µm	
17.0000 mm		17.0012 mm		1.2000 µm		± 4 µm		± 6.6E+01 µm	
20.0000 mm		19.9978 mm		-2.2000 µm		± 4 µm		± 7.8E+01 µm	
22.0000 mm		21.9998 mm		-0.2000 µm		± 4 µm		± 8.5E+01 µm	
25.0000 mm		25.0005 mm		0.5000 µm		± 4 µm		± 4.0E+00 µm	