

Certificado No. (Certificate #):		57708		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-06-11		Marca (Brand):		Mitutoyo	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		10 mm	9.95 mm	-50 µm	± 50 µm	± 7.6E+01 µm	
Exterior / (Exterior)		30 mm	29.95 mm	-50 µm	± 50 µm	± 7.6E+01 µm	
Exterior / (Exterior)		100 mm	99.93 mm	-67 µm	± 50 µm	± 7.6E+01 µm	
Exterior / (Exterior)		200 mm	199.90 mm	-100 µm	± 70 µm	± 7.6E+01 µm	
Exterior / (Exterior)		280 mm	279.88 mm	-117 µm	± 80 µm	± 7.6E+01 µm	
Interior / (Interior)		50 mm	50.00 mm	0 µm	± 50 µm	± 4.1E+01 µm	
Interior / (Interior)		150 mm	150.05 mm	50 µm	± 50 µm	± 4.1E+01 µm	