

Certificado No. (Certificate #):		57709		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-06-09		Marca (Brand):		NEIKO	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		1 mm	1.00 mm	0 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		5 mm	5.00 mm	0 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		20 mm	20.00 mm	0 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		50 mm	50.00 mm	0 µm	± 30 µm	± 7.4E+00 µm	
Exterior / (Exterior)		130 mm	129.98 mm	-20 µm	± 30 µm	± 7.5E+00 µm	
Interior / (Interior)		50 mm	49.97 mm	-30 µm	± 30 µm	± 8.2E+00 µm	
Interior / (Interior)		150 mm	149.92 mm	-77 µm	± 30 µm	± 8.2E+00 µm	
Profundidad / (Depth)		30 mm	30.02 mm	20 µm	± 30 µm	± 7.1E+00 µm	
Profundidad / (Depth)		100 mm	100.02 mm	23 µm	± 30 µm	± 7.1E+00 µm	