

Certificado No. (Certificate #):57711Equipo (Instrument):Reloj Indicador

Fecha de Calibración (Calibration Date):2025-06-12Marca (Brand):Gilson

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
0.0500 in	0.051 in	0.001 in	± 0.001 in	±	8.2E-04 in
0.1000 in	0.100 in	0.000 in	± 0.001 in	±	5.8E-04 in
0.1050 in	0.105 in	0.000 in	± 0.001 in	±	5.8E-04 in
0.1500 in	0.151 in	0.001 in	± 0.001 in	±	8.2E-04 in
0.2000 in	0.200 in	0.000 in	± 0.001 in	±	5.9E-04 in
0.3000 in	0.300 in	0.000 in	± 0.001 in	±	6.1E-04 in
0.4000 in	0.400 in	0.000 in	± 0.001 in	±	6.4E-04 in
0.5000 in	0.500 in	0.000 in	± 0.001 in	±	6.8E-04 in
1.0000 in	1.000 in	0.000 in	± 0.001 in	±	9.1E-04 in