

Certificado No. (Certificate #):		57712		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-06-09		Marca (Brand):		NEIKO	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		10 mm	10.01 mm	7 µm	± 30 µm	± 7.7E+00 µm	
Exterior / (Exterior)		50 mm	49.99 mm	-7 µm	± 30 µm	± 7.7E+00 µm	
Exterior / (Exterior)		130 mm	129.97 mm	-33 µm	± 30 µm	± 7.7E+00 µm	
Exterior / (Exterior)		200 mm	199.96 mm	-40 µm	± 30 µm	± 7.7E+00 µm	
Exterior / (Exterior)		280 mm	279.96 mm	-40 µm	± 40 µm	± 7.7E+00 µm	
Interior / (Interior)		50 mm	49.99 mm	-10 µm	± 30 µm	± 8.2E+00 µm	
Interior / (Interior)		150 mm	149.98 mm	-20 µm	± 30 µm	± 8.2E+00 µm	
Profundidad / (Depth)		30 mm	30.02 mm	20 µm	± 30 µm	± 7.1E+00 µm	
Profundidad / (Depth)		100 mm	100.03 mm	27 µm	± 30 µm	± 7.1E+00 µm	