

Certificado No. (Certificate #):		57726		Equipo (Instrument):		Calibrador Pie de Rey	
Fecha de Calibración (Calibration Date):		2025-06-17		Marca (Brand):		HARDEN	
Tipo de Medida (Measurement Type)		Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
Exterior / (Exterior)		10 mm	10.00 mm	0 µm	± 50 µm	± 4.0E+01 µm	
Exterior / (Exterior)		25 mm	25.00 mm	0 µm	± 50 µm	± 4.0E+01 µm	
Exterior / (Exterior)		80 mm	80.00 mm	0 µm	± 50 µm	± 4.0E+01 µm	
Exterior / (Exterior)		120 mm	120.00 mm	0 µm	± 50 µm	± 4.0E+01 µm	
Exterior / (Exterior)		200 mm	200.00 mm	0 µm	± 70 µm	± 4.0E+01 µm	
Interior / (Interior)		130 mm	130.00 mm	0 µm	± 50 µm	± 3.1E+01 µm	
Interior / (Interior)		150 mm	150.00 mm	0 µm	± 50 µm	± 3.1E+01 µm	
Profundidad / (Depth)		50 mm	50.00 mm	0 µm	± 50 µm	± 3.3E+01 µm	
Profundidad / (Depth)		100 mm	100.00 mm	0 µm	± 50 µm	± 3.3E+01 µm	